



## CERTIFICATE OF ANALYSIS

Stat Profile pHox/Basic Calibrator Cartridge C  
Lot Number 24150028

Ref Number 23936  
Exp Date 2025-10-29  
Date of Manufacture 2024-04-29  
UDI# (01)00385480239363(11)240429(17)251029(10)24150028

Cal A L/N 24123026			Cal B L/N 24123043			Cal C L/N 24122049		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HCO <sub>3</sub> mmol/L	20.1 - 20.9	Pass	HCO <sub>3</sub> mmol/L	40.2 - 41.8	Pass	pH @ 37C	7.342 - 7.348	Pass
Na mmol/L	139.3 - 141.3	Pass						
Cal D L/N 24122009			Reference L/N 24120060					
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>			
pH @ 37C	6.792 - 6.798	Pass	KCl mol/L	1.8 - 2.2	Pass			
Na mmol/L	71.5 - 72.5	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi M*

Approval

QC Inspector  
Title

06/3/2024  
Date