



CERTIFICATE OF ANALYSIS

Bio pHox Cal Pack
Lot Number 24150029

Part Number 41874
Exp Date 2025-10

Std A	L/N	24123026									
			<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>	
Na mmol/L			139.0 - 141.0	Pass	HCO3 mmol/L	39.2 - 40.8	Pass	pH @ 37C	7.352 - 7.358	Pass	
HCO3 mmol/L			19.6 - 20.4	Pass							
	Std D	L/N	24122009			Reference	L/N	24120060			
			<u>Spec</u>	<u>Pass/Fail</u>				<u>Spec</u>	<u>Pass/Fail</u>		
	pH @ 37C		6.802 - 6.808	Pass		KCl mol/L		1.8 - 2.2	Pass		
	Na mmol/L		69.5 - 70.5	Pass							

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCl.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

6/3/2024
Date