



**CERTIFICATE OF ANALYSIS**

Nova Glu/Lac Sensor Soaking Solution  
Lot Number 24123011

Ref Number 13409

Exp Date 2025-05-02

Date of Manufacture 2024-05-02

UDI# (01)00385480134095(11)240502(17)250502(10)24123011

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.43 - 0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

*Awishi M*

Approval

QC Inspector

Title

6/5/2024

Date