



CERTIFICATE OF ANALYSIS

Nova Primary Linearity Level 2

Ref Number 63166

Lot Number 24008045

Exp Date 2025-07-08

Date of Manufacture 2024-01-08

UDI# (01)10385480631669(11)240108(17)250708(10)24008045

Level 2

	<u>Spec</u>	<u>Pass/Fail</u>
Glu mg/dL	655-739	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

Quality Control Manager

Title

6/5/2024

Date