



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	<u>PASS</u> / FAIL <u>PASS</u> / FAIL <u>PASS</u> / FAIL
Qty:	6,480
Lot No.	4924149129
Mfg. Date:	2024 - 05 - 28
Expiration Date:	2025 - 05 - 28
Unique Device Identifier (UDI):	(01)20385480432720 (11) <u>240528</u> (17) <u>250528</u> (10) <u>4924149129</u>

Marc Brosme
Quality Control Manager

6/7/2024
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454