



CERTIFICATE OF ANALYSIS

Preheater Deproteinizing Sol'n
Lot Number 22299061

Ref Number 12704
Exp Date 2023-10-26
Date of Manufacture 2022-10-26
UDI# (01)00385480127042(11)221026(17)231026(10)22299061

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.70 - 0.90%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

11/16/2022
Date