



CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamine Membrane
Lot Number 24156002

Part Number 42486
Exp Date 2024-12

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	33.16
Biopro Ctl L4	0.74 - 1.24	0.99
Biopro Ctl L5	4.43 - 5.93	5.21

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

6/10/2024
Date