



CERTIFICATE OF ANALYSIS

Prime CCS Comp Cal Cartridge VET 400 Sample
Lot Number 24143038

Ref Number 53106
Exp Date 2025-11-16
Date of Manufacture 2024-05-16

Cal A L/N 24141029			Cal B L/N 24137015		
	<u>Assigned Value*</u>	<u>Result</u>		<u>Spec</u>	<u>Pass/Fail</u>
pH	7.355	N/A	pH	6.837 - 6.843	Pass
pCO2 mmHg	24.8	N/A	pO2 mmHg	167.0	N/A
			Na mmol/L	68.0 - 69.0	Pass
			K mmol/L	9.92 - 10.08	Pass
			Cl mmol/L	45.6 - 46.4	Pass
			iCa mmol/L	2.17 - 2.23	Pass
			Glu mg/dL	197 - 203	Pass
			Lac mmol/L	9.5 - 10.5	Pass
	<u>Spec</u>	<u>Pass/Fail</u>			
Na mmol/L	132.1 - 133.9	Pass			
K mmol/L	3.97 - 4.03	Pass			
Cl mmol/L	105.2 - 106.8	Pass			
iCa mmol/L	1.08 - 1.12	Pass			
Glu mg/dL	79 - 81	Pass			
Lac mmol/L	1.8 - 2.2	Pass			

Flush L/N 24137013			Reference L/N 24137016		
	<u>Assigned Value*</u>	<u>Result</u>		<u>Spec</u>	<u>Pass/Fail</u>
pCO2 mmHg	51.2	N/A	KCl	1.8 - 2.2	Pass
pO2 mmHg	93.1	N/A			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

06/11/2024
Date