



CERTIFICATE OF ANALYSIS

Nova 4/10/11/13 Lithium Sensor
Lot Number 24144038

Ref Number 15742
Exp Date 2025-05-23

Date of Manufacture 2024-05-23

UDI# (01)10385480157428(11)240523(17)250523(10)24144038

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	10.3 - 11.7	Pass
Chem Ctl L 2 mmol/L	1.77 - 2.27	Pass
Set A L 2 mmol/L	1.96 - 2.36	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

²Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

6/12/2024

Date