

CERTIFICATE OF ANALYSIS

Preheater Deproteinizing Sol'n Lot Number 22357021

Ref Number 12704

1 Exp Date 2023-12-23

Date of Manufacture 2022-12-23

UDI# (01)00385480127042(11)221223(17)231223(10)22357021

Spec Pass/Fail
NaOCI 0.70 - 0.90% Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector Title

1/13/2023 Date