



**CERTIFICATE OF ANALYSIS**

Nova Lactate Plus Control Solution Level 2  
Lot Number 3824137302

Ref Number 40814  
Exp Date 2026-11-16  
Date of Manufacture 2024-05-16  
UDI# (01)00385480408141(11)240516(17)261116(10)3824137302

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	4.4 - 5.0	4.5

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi M*

Approval

QC Inspector

Title

6/14/2024

Date