



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample
Lot Number 24163036

Ref Number 53463
Exp Date 2025-12-03
Date of Manufacture 2024-06-03
UDI# (01)00385480534635(11)240603(17)251203(10)24163036

| | | | |
|-------|-----------|------------------------|---------------|
| Cal A | L/N | 24155037 | |
| | | <u>Assigned Value*</u> | <u>Result</u> |
| | pH | 7.352 | N/A |
| | pCO2 mmHg | 25.0 | N/A |

| | | | |
|-------|----------|---------------|------------------|
| Cal B | L/N | 24155038 | |
| | | <u>Spec</u> | <u>Pass/Fail</u> |
| | pH | 6.837 - 6.843 | Pass |
| | pO2 mmHg | 167.0 | N/A |

| | | | |
|-------|-----------|------------------------|---------------|
| Flush | L/N | 24155024 | |
| | | <u>Assigned Value*</u> | <u>Result</u> |
| | pCO2 mmHg | 50.4 | N/A |
| | pO2 mmHg | 91.3 | N/A |

| | | | |
|-----------|-----|-------------|------------------|
| Reference | L/N | 24156013 | |
| | | <u>Spec</u> | <u>Pass/Fail</u> |
| | KCl | 1.8 - 2.2 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

06/18/2024
Date