



**CERTIFICATE OF ANALYSIS**

Nova Control Level 3  
Lot Number 24129074

Ref Number 18654  
Level 3 Exp Date 2026-05-08  
Date of Manufacture 2024-05-08  
UDI# (01)00385480186544(11)240508(17)260508(10)24129074

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ mmol/L	145-171	Pass
K+ mmol/L	1.7-2.3	Pass
Cl- mmol/L	105-127	Pass
TCO2 mmol/L	28-40	Pass
Glu mg/dL	312-408	Pass
BUN mg/dL	21-35	Pass
TCa mg/dL	2.68-4.31	Pass
iCa++ mg/dL	2.08-3.31	Pass
iMg++ mg/dL	0.74-1.23	Pass
Li+ mmol/L	2.59-3.50	Pass
pH	7.25-7.50	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

*Awish M*

Approval

QC Inspector

Title

6/18/2024

Date