



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 500 Sample  
Lot Number 24163036

Ref Number 53464  
Exp Date 2025-12-03  
Date of Manufacture 2024-06-03  
UDI# (01)00385480534642(11)240603(17)251203(10)24163036

Cal A	L/N	24155037		
		<u>Assigned Value*</u>	<u>Result</u>	
pH		7.352	N/A	
pCO2 mmHg		25.0	N/A	

Cal B	L/N	24155038		
		<u>Spec</u>	<u>Pass/Fail</u>	
pH		6.837 - 6.843	Pass	
pO2 mmHg		167.0	N/A	

Flush	L/N	24155024		
		<u>Assigned Value*</u>	<u>Result</u>	
pCO2 mmHg		50.4	N/A	
pO2 mmHg		91.3	N/A	

Reference	L/N	24156013		
		<u>Spec</u>	<u>Pass/Fail</u>	
KCl		1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

06/18/2024  
Date