



**CERTIFICATE OF ANALYSIS**

BioPro Flex Cell Density/Viability Reagent Pack  
Lot Number 24169021

Ref Number 39581  
Exp Date 2025-05

W Soln L/N	24150032		C Soln L/N	24152022	
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
Depro	N/A	Present	3% Detergent pH	N/A >9.0	Present Pass
A Soln L/N	24157011		T Soln L/N	24150016	
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
63% Ethanol	N/A	Present	0.4% Trypan	N/A	Present
3.5% Methanol	N/A	Present	Particle Count	≤ 2.0	Pass
3.5% Isopropanol	N/A	Present			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Haneen Aldarwish*

Approval

QC Inspector  
Title

6/19/2024  
Date