OI-922

Page 40 of 66

Rev. BJ

Vialed and Packaged Test Strip Release

Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:		LIFESCAN
Description:		TEST STRIPS, STATSTRIP TM GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT
Lifescan Purchase Order No.:		478756
Catalog Number:		42214
Nova Biomedical P/N		54311
One Touch P/N		02307501
Target Range (mg/dL) SD(mg/dL) /CV (%)		
30-50	≤4 mg/dL	PASS FAIL
51-110	≤ 6.0%	PASS FAIL
111-150	≤ 5.0%	PASS FAIL
151-250	≤ 5.0%	PASS FAIL
251-400	≤ 5.0%	PASS FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method \geq 95%		PASS FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%		PASS / FAIL
Qty:		18,144
Lot No.		0324162309
Date of Mfg. (DOM):		2024/06/10 (YYYY/MM/DD)
Expiration Date:		2026/12 (YYYY/MM)
Unique Device Identifier (UDI):		(01)20385480422141
		(11)2 4 0 6 1 0
		(17) 2 6 1 1 1 1 1 1 1 1 1 1
01 . 0		(10)0324162309

Nova Biomedical Corporation 200 Prospect St

Waltham, MA 02454