



**CERTIFICATE OF ANALYSIS**

Bioprofile Flex Lactate Membrane  
Lot Number 24177011

Part Number 42482  
Exp Date 2025-06

|               | <u>Spec*</u>  | <u>Mean Result</u> |
|---------------|---------------|--------------------|
| Slope         | 10.00 - 60.00 | 17.09              |
| Biopro Ctl L1 | 0.78 - 1.28   | 0.99               |
| Biopro Ctl L3 | 4.25 - 5.75   | 5.16               |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

7/8/2024  
Date