



CERTIFICATE OF ANALYSIS

Bio pHox Calibrator Cartridge
Lot Number 24180059

Ref Number 41874
Exp Date 2025-12

Cal A L/N 24176051

Cal B L/N 24177035

Cal C L/N 24176052

	<u>Spec</u>	<u>Pass/Fail</u>
HCO3 mmol/L	139.0 - 141.0	Pass
Na mmol/L	19.6 - 20.4	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
HCO3 mmol/L	39.2 - 40.8	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	7.352 - 7.358	Pass

Cal D L/N 24177071

Reference L/N 24176021

	<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	6.802 - 6.808	Pass
Na mmol/L	69.5 - 70.5	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
KCl mol/L	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

07/8/2024
Date