



CERTIFICATE OF ANALYSIS

Nova 1-11, 13 Dilutor Pack
Lot Number 22306011

Ref Number 12882
Exp Date 2023-11-02
Date of Formulation 2022-11-02
UDI# (01)00385480128827(11)221102(17)231102(10)22306011

Cal L L/N 22306011

| | | <u>Spec</u> | <u>Pass/Fail</u> |
|----|--------|-------------|------------------|
| Na | mmol/L | 29.8 - 30.2 | Pass |
| K | mmol/L | 9.9 - 10.1 | Pass |
| Cl | mmol/L | 39.7 - 40.3 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

1/14/2023

Date