

CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge Lot Number 24183017 Ref Number 54652 Exp Date 2026-01-01 Date of Manufacture 2024-07-01 UDI# (01)00385480546522(11)240701(17)260101(10)24183017

Ctl Level 1				Ctl Level 2		
	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail	
HDL mg/dL	28 - 50	Pass	HDL mg/dL	68 - 102	Pass	
TCHOL mg/dL	155 - 225	Pass	TCHOL mg/dL	307 - 423	Pass	
TGL mg/dL	60 - 136	Pass	TGL mg/dL	280 - 420	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/9/2024 Date