

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, NOVA PRO GLUCOSE, PACKAGED 1800 COUNT
Catalog Number:	53881
Nova Biomedical P/N	53881
Target Range (mg/dL) SD(mg/dL) /CV (%)	
30-50 ≤ 4 mg/dL	PASS / FAIL
51-110 ≤ 6.0%	PASS / FAIL
111-150 ≤ 5.0%	PASS / FAIL
151-250 ≤ 5.0%	PASS / FAIL
251-400 ≤ 5.0%	PASS / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method ≥ 95%	PASS / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method ≥ 98%	PASS / FAIL
Qty:	1,890
Lot No.	1890224169
Mfg. Date:	2024-06-17
Expiration Date:	2026-06-17
Unique Device Identifier (UDI):	(01)20385480538811 (11) <u>240617</u> (17) <u>260617</u> (10) <u>1890224169</u>

Marc Broome
 Quality Control Manager

6/28/2024
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454