



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 500 Sample  
Lot Number 24171069

Ref Number 53464  
Exp Date 2025-12-11  
Date of Manufacture 2024-06-11  
UDI# (01)00385480534642(11)240611(17)251211(10)24171069

Cal A	L/N	24164078	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.351	N/A
	pCO2 mmHg	25.1	N/A

Cal B	L/N	24164079	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	24163026	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	50.8	N/A
	pO2 mmHg	94.9	N/A

Reference	L/N	24163095	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

07/16/2024  
Date