



CERTIFICATE OF ANALYSIS

Nova 1 Calibrator Pack
Lot Number 416629

Ref Number 15646
Exp Date 2026-06-14
Date of Manufacture 2024-06-14
UDI# (01)00385480156462(11)240614(17)260614(10)416629

Cal A L/N	24166029	Cal B L/N	24166029	Reference L/N	24166029			
	<u>Spec</u>	<u>Pass/Fail</u>	<u>Spec</u>	<u>Pass/Fail</u>	<u>Spec</u>	<u>Pass/Fail</u>		
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	59.4 - 60.6	Pass	KCl mol/L	1.8 - 2.2	Pass
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

7/15/2024
Date