



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ pH Sensor
Lot Number 24172054

Part Number 48580
Exp Date 2025-12-20
Date of Manufacture 2024-06-20

UDI# (01)00385480485807(11)240620(17)251220(10)24172054

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.20 - 11.20	Pass
Std C mV	0.00 ± 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/18/2024
Date