

CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge Lot Number 24190019 Ref Number 54652 Exp Date 2026-01-08 Date of Manufacture 2024-07-08 UDI# (01)00385480546522(11)240708(17)260108(10)24190019

Ctl Level 1

Ctl Level 2

	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail
HDL mg/dL	28 - 50	Pass	HDL mg/dL	68 - 102	Pass
TCHOL mg/dL TGL mg/dL	155 - 255 60 - 136	Pass Pass	TCHOL mg/dL TGL mg/dL	307 - 423 280 - 420	Pass Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Frut

Approval

QC Inspector Title

7/18/2024 Date