



CERTIFICATE OF ANALYSIS

Nova StatStrip Hb/Hct Control Level 2
Lot Number 2024159242

Ref Number 59827
Exp Date 2026-06-07

Date of Manufacture 2024-06-07

UDI# (01)00385480598279(11)240607(17)260607(10)2024159242

	<u>Spec</u>	<u>Mean Result</u>
Hb g/dL	13.4 - 14.6	14.4
Hct %	38 - 42	41

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

7/17/2024
Date