



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 24180007

Part Number 48554
Exp Date 2025-09-28
Date of Manufacture 2024-06-28
UDI# (01)00385480485548(11)240628(17)250928(10)24180007

| | <u>Spec*</u> | <u>Mean Result</u> |
|------------------------|--------------|--------------------|
| Slope | ≥ 15.00 | 37.22 |
| pHox Ultra/CCX Ctl L4 | 1.7 - 2.3 | 2.0 |
| pHox Ultra/CCX Ctl L5 | 6.3 - 7.7 | 7.4 |
| Lin Soln Set A Level 1 | 14.0 - 20.0 | 18.8 |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/19/2024
Date