

CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamine Membrane Lot Number 24192014

Part Number 42486 Exp Date 2025-01

| | <u>Spec*</u> | Mean Result |
|---------------|---------------|-------------|
| Slope | 10.00 - 60.00 | 25.36 |
| Biopro Ctl L4 | 0.74 - 1.24 | 0.96 |
| Biopro Ctl L5 | 4.43 - 5.93 | 4.98 |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Stat QuiM

Approval

QC Inspector Title

7/19/2024 Date