



CERTIFICATE OF ANALYSIS

Bioprofile pH Linearity Solution LOW
Lot Number 24159036

Catalog Number 49939
Exp Date 2025-06

Spec	Pass/Fail
pH 6.200 - 6.400	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials:
pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/19/2024
Date