



CERTIFICATE OF ANALYSIS

PCO₂ / TCO₂ CONDITIONING SOLUTION
Lot Number 24185042

Part Number 06857
Exp Date 2025-10-03
Date of Manufacturing 2024-07-03
UDI# (01)00385480068574(11)240703(17)251003(10)24185042

	<u>Spec</u>	<u>Pass/Fail</u>
pH	0.70 - 1.30	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Haneen Aldarwish

Approval

QC Inspector
Title

7/24/2024
Date