



CERTIFICATE OF ANALYSIS

Nova Glu/Lac Sensor Soaking Solution
Lot Number 22257037

Ref Number 13409

Exp Date 2023-09-14

Date of Manufacture 2022-09-14

UDI# (01)00385480134095(11)220914(17)230914(10)22257037

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.43 - 0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Awish M

Approval

QC Inspector

Title

10/6/2022

Date