



CERTIFICATE OF ANALYSIS

Preheater Deproteinizing Sol'n
Lot Number 24183077

Ref Number 12704
Exp Date 2025-07-01
Date of Manufacture 2024-07-01
UDI# (01)00385480127042(11)240701(17)250701(10)24183077

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.70 - 0.90%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

7/25/2024
Date