OI-922 Page

Page 39 of 66

Rev. BK

Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	LIFESCAN
Description:	TEST STRIPS, STATSTRIP <sup>™</sup> GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT
Lifescan Purchase Order No.:	912411
Catalog Number:	42214
Nova Biomedical P/N	54515
One Touch P/N	02315302
Target Range (mg/dL) SD(mg/dL) /C	V (%)
30-50 ≤ 4 mg/	/dL PASS FAIL
51 <b>-</b> 110 ≤ 6.0%	PASS FAIL
111-150 $\leq 5.0\%$	PASS / FAIL
$151-250 \leq 5.0\%$	PASS FAIL
$251-400 \leq 5.0\%$	PASS FAIL
Accuracy Criteria 1: Total within +/-12 mg/ 12% of reference method $\geq$ 95%	/dL or +/- PASS FAIL
Accuracy Criteria 2: Total within +/-15 mg/ 15% of reference method ≥ 98%	/dL or +/-
Qty:	30,240
Lot No.	0324197249
Date of Mfg. (DOM):	2024 / 07/15 (YYYY/MM/DD)
Expiration Date:	2026/07/16 (YYYY/MM/DD)
Unique Device Identifier (UDI):	(01)20385480422141
	$(11)_{2}$ $4$ $0$ $7$ $1$ $\xi$
	$(17)_{2}_{6}_{0}_{7}_{1}_{1}_{5}$
. /	(10)1) 3 2 4 1 4 1 2 4 1

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect St

Waltham, MA 02454