



CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge
Lot Number 24206022

Ref Number 54215
Exp Date 2025-07-24
Date of Manufacture 2024-07-24
UDI# (01)00385480542159(11)240724(17)250724(10)24206022

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HbA1c %	4.9 - 6.7	Pass	HbA1c %	8.1 - 11.7	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/29/2024

Date