



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 24204038

Ref Number 53359
Exp Date 2026-01-02
Date of Manufacture 2024-07-02
UDI# (01)00385480533591(11)240702(17)260102(10)24204038

Cal A	L/N	24184012		Cal B	L/N	24184013	
		<u>Assigned Value*</u>	<u>Result</u>			<u>Spec</u>	<u>Pass/Fail</u>
	pH	7.350	N/A		pH	6.837 - 6.843	Pass
pCO2	mmHg	24.9	N/A	pO2	mmHg	167.0	N/A
Flush	L/N	24192073		Reference	L/N	24184014	
		<u>Assigned Value*</u>	<u>Result</u>			<u>Spec</u>	<u>Pass/Fail</u>
pCO2	mmHg	50.9	N/A	KCl		1.8 - 2.2	Pass
pO2	mmHg	95.6	N/A				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

07/30/2024
Date