

OI-922

Page 61 of 66

Rev. BK

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

| Customer: | N/A | | | | | | | | | | | | | | | |
|--|--|--------------|-----------|----------------|----------------|--------------|----------------|---------|-------|----------------|---------|-------|--|-------------|-------------|-------------|
| Description: | TEST STRIPS, NOVA MAX KETONE, PACKAGED | | | | | | | | | | | | | | | |
| Catalog Number: | 53493 | | | | | | | | | | | | | | | |
| Nova Biomedical P/N | 53493 | | | | | | | | | | | | | | | |
| <table border="1"><thead><tr><th>Target Range</th><th>MPE</th><th>SD/CV (%)</th></tr></thead><tbody><tr><td>0.1-0.8 mmol/L</td><td>+/- 0.3 mmol/L</td><td>≤ 0.3 mmol/L</td></tr><tr><td>1.5-2.5 mmol/L</td><td>+/- 20%</td><td>≤ 15%</td></tr><tr><td>3.5-5.0 mmol/L</td><td>+/- 20%</td><td>≤ 15%</td></tr></tbody></table> | Target Range | MPE | SD/CV (%) | 0.1-0.8 mmol/L | +/- 0.3 mmol/L | ≤ 0.3 mmol/L | 1.5-2.5 mmol/L | +/- 20% | ≤ 15% | 3.5-5.0 mmol/L | +/- 20% | ≤ 15% | <table border="1"><tbody><tr><td>PASS / FAIL</td></tr><tr><td>PASS / FAIL</td></tr><tr><td>PASS / FAIL</td></tr></tbody></table> | PASS / FAIL | PASS / FAIL | PASS / FAIL |
| Target Range | MPE | SD/CV (%) | | | | | | | | | | | | | | |
| 0.1-0.8 mmol/L | +/- 0.3 mmol/L | ≤ 0.3 mmol/L | | | | | | | | | | | | | | |
| 1.5-2.5 mmol/L | +/- 20% | ≤ 15% | | | | | | | | | | | | | | |
| 3.5-5.0 mmol/L | +/- 20% | ≤ 15% | | | | | | | | | | | | | | |
| PASS / FAIL | | | | | | | | | | | | | | | | |
| PASS / FAIL | | | | | | | | | | | | | | | | |
| PASS / FAIL | | | | | | | | | | | | | | | | |
| Qty: | 5,310 | | | | | | | | | | | | | | | |
| Lot No.: | 1510224197 | | | | | | | | | | | | | | | |
| Mfg. Date: | 2024-07-15 | | | | | | | | | | | | | | | |
| Expiration Date: | 2026-07-15 | | | | | | | | | | | | | | | |
| Unique Device Identifier (UDI): | (01)20385480 53493 6 (11)240715 (17)260715 (10)1510224197 | | | | | | | | | | | | | | | |

Marc Broome
Quality Control Manager

8/5/24
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454