

## CERTIFICATE OF ANALYSIS

Bio pHOx Cal Pack Lot Number 24215031 Part Number 41874 Exp Date 2026-01

Std A L/N 24207026	Std B L/N 24206016		Std C L/N 24205037		
<u>Spec</u> Na mmol/L 139.0 - 141.0 HCO3 mmol/L 19.6 - 20.4	<u>Pass/Fail</u> Pass Pass	<u>Sper</u> HCO3 mmol/L 39.2 - 4		pH @ 37C	<u>Spec Pass/Fail</u> ; 7.352 - 7.358 Pass
Std D L/N	24205038		Reference L/N	24205039	
pH @ 37C Na mmol/L	<u>Spec</u> 6.802 - 6.808 69.5 - 70.5	<u>Pass/Fail</u> Pass Pass	KCI mol/L	<u>Spec</u> 1.8 - 2.2	<u>Pass/Fail</u> Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCI.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Amishi.M

Approval

QC Inspector Title

<u>8/6/2024</u> Date