



**CERTIFICATE OF ANALYSIS**

Bioprofile Flex Glutamine Membrane  
Lot Number 24207001

Part Number 42486  
Exp Date 2025-01

|               | <u>Spec*</u>  | <u>Mean Result</u> |
|---------------|---------------|--------------------|
| Slope         | 10.00 - 60.00 | 26.11              |
| Biopro Ctl L4 | 0.74 - 1.24   | 0.99               |
| Biopro Ctl L5 | 4.37 - 5.87   | 5.11               |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

8/5/2024  
Date