



CERTIFICATE OF ANALYSIS

Allegro Control 1 UACR
Lot Number 8524197241
Ref Number 51032
Exp Date 2026-07-15
Date of Manufacture 2024-07-15
UDI# (01)00385480510325(11)240715(17)260715(10)8524197241

| | Spec | Pass/Fail |
|-------------------|--------|-----------|
| Albumin, mg/dL | 33-81 | Pass |
| Creatinine, mg/dL | 65-119 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/7/2024
Date