



CERTIFICATE OF ANALYSIS

Prime CCS Comp Cal Cartridge VET 400 Sample
Lot Number 24212067

Ref Number 53106
Exp Date 2026-01-16
Date of Manufacture 2024-07-16

Cal A L/N	24198051			Cal B L/N	24198052		
		<u>Assigned Value*</u>	<u>Result</u>			<u>Spec</u>	<u>Pass/Fail</u>
pH		7.354	N/A	pH		6.837 - 6.843	Pass
pCO2 mmHg		24.6	N/A	pO2 mmHg		167.0	N/A
		<u>Spec</u>	<u>Pass/Fail</u>	Na mmol/L		68.0 - 69.0	Pass
Na mmol/L		132.1 - 133.9	Pass	K mmol/L		9.92 - 10.08	Pass
K mmol/L		3.97 - 4.03	Pass	Cl mmol/L		45.6 - 46.4	Pass
Cl mmol/L		105.2 - 106.8	Pass	iCa mmol/L		2.17 - 2.23	Pass
iCa mmol/L		1.08 - 1.12	Pass	Glu mg/dL		197 - 203	Pass
Glu mg/dL		79 - 81	Pass	Lac mmol/L		9.5 - 10.5	Pass
Lac mmol/L		1.8 - 2.2	Pass				

Flush L/N	24198049			Reference L/N	24204045		
		<u>Assigned Value*</u>	<u>Result</u>			<u>Spec</u>	<u>Pass/Fail</u>
pCO2 mmHg		51.3	N/A	KCl		1.8 - 2.2	Pass
pO2 mmHg		98.0	N/A				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

08/07/2024
Date