



CERTIFICATE OF ANALYSIS

Linearity Standard Set D Level 1
Lot Number 20059016

Ref Number 13523
Exp Date 2023-02-28
Date of Manufacture 2020-02-28
UDI# (01)00385480135238(11)200228(17)230228(10)20059016

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ mmol/L	108 - 118	Pass
Hct %	13 - 17	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02453 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Inspector
Title

03-10-2020
Date