



CERTIFICATE OF ANALYSIS

Linearity Standard Set D Level 1

Lot Number 22131015

Ref Number 13523

Exp Date 2025-05-11

Date of Manufacture 2022-05-11

UDI# (01)00385480135238(11)220511(17)250511(10)22131015

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ mmol/L	108 - 118	Pass
Hct %	13 - 17	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

6/10/2022

Date