



CERTIFICATE OF ANALYSIS

Nova 12/14/16 Dilutor Pack
Lot Number 24214028

Ref Number 15819
Exp Date 2025-08-01
Date Of Manufacture 2024-08-01
UDI# (01)00385480158190(11)240801(17)250801(10)24214028

Cal L L/N	24214028	
	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	29.8 - 30.2	Pass
K mmol/L	9.9 - 10.1	Pass
Cl mmol/L	39.7 - 40.3	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Haneen Aldarwish

Approval

QC Inspector
Title

8/12/2024
Date