OI-922

Page 40 of 66

Rev. BK

Vialed and Packaged Test Strip Release

Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	LIFESCAN
Description:	TEST STRIPS, STATSTRIP TM GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT
Lifescan Purchase Order No.:	481269
Catalog Number:	42214
Nova Biomedical P/N	54311
One Touch P/N	02307501
Target Range (mg/dL) SD(mg/dL) /CV (%)	
30-50 ≤4 mg/dL	PASS FAIL
51-110 ≤ 6.0%	PASS/ FAIL
111-150 ≤ 5.0%	PASS FAIL
$151-250 \leq 5.0\%$	PASS / FAIL
$251-400 \leq 5.0\%$	PASS / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method \geq 95%	PASS FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method \geq 98%	PASS / FAIL
Qty:	21,168
Lot No.	0324207309
Date of Mfg. (DOM):	2024 / 07 / 25 (YYYY/MM/DD)
Expiration Date:	2027 / 01 (YYYY/MM)
Unique Device Identifier (UDI):	$\begin{array}{c} (01)20385480422141 \\ (11)\underline{2} \ \underline{4} \ \underline{0} \ \underline{7} \ \underline{2} \ \underline{5} \\ (17) \ \underline{2} \ \underline{7} \ \underline{0} \ \underline{1} \ \underline{2} \ \underline{5} \\ (10) \ \underline{0} \ \underline{3} \ \underline{2} \ \underline{11} \ \underline{2} \ \underline{0} \ \underline{7} \ \underline{3} \ \underline{0} \ \underline{9} \end{array}$

Quality Control Manager

8/13/24

Date

Nova Biomedical Corporation

200 Prospect St

Waltham, MA 02454