



CERTIFICATE OF ANALYSIS

Bioprofile Flex Lactate Membrane
Lot Number 24222027

Part Number 42482
Exp Date 2025-08

| | <u>Spec*</u> | <u>Mean Result</u> |
|---------------|---------------|--------------------|
| Slope | 10.00 - 60.00 | 19.75 |
| Biopro Ctl L1 | 0.78 - 1.28 | 0.96 |
| Biopro Ctl L3 | 4.25 - 5.75 | 5.08 |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/13/2024
Date