OI-922

Page 33 of 66

Rev. BK

Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:		
customer.		NIA
Description:		TEST STRIPS, STATSTRIP TM GLUCOSE, PACKAGED 1800 COUNT
Catalog Number:		42214
Nova Biomedical P/N		51493
Target Range (mg/dL) SD	(mg/dL) /CV (%)	31413
30-50	$\leq$ 4 mg/dL	PASS FAIL
51-110	≤ 6.0%	PASS FAIL
111-150	≤ 5.0%	PASS FAIL
151-250	≤ 5.0%	PASS FAIL
251-400	≤ 5.0%	PASS FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method $\geq$ 95%		PASS FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- $15\%$ of reference method $\geq 98\%$		PASS FAIL
Qty:		1,470
Lot No.		0324207309
Mfg. Date:		2024-07-25
Expiration Date:		2027-01-25
Unique Device Identifier (UDI):		(01)20385480422141
		(11) 2 4 0 7 2 5
		(17) $2$ $7$ $0$ $1$ $2$ $5$ $10$ $0$ $3$ $1$ $4$ $2$ $0$ $7$ $3$ $0$ $9$
March	ane	8 14 2024

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454