




### CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	<del>PASS/FAIL</del> <del>PASS/FAIL</del> <del>PASS/FAIL</del>
Qty:	7,452
Lot No.	4924218129
Mfg. Date:	2024-08-05
Expiration Date:	2025-08-05
Unique Device Identifier (UDI):	(01)20385480432720 (11)240805 (17)250805 (10)4924218129

  
Quality Control Manager

8/19/24  
Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454