



**CERTIFICATE OF ANALYSIS**

Bioprofile Flex Glutamine Membrane  
Lot Number 24225013

Part Number 42486  
Exp Date 2025-02

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	28.74
Biopro Ctl L4	0.74 - 1.24	0.94
Biopro Ctl L5	4.37 - 5.87	4.88

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

8/19/2024  
Date