



CERTIFICATE OF ANALYSIS

Nova 1/4/5/8/10/11/12/13/14 Potassium Sensor
Lot Number 24225039

Ref Number 14713
Exp Date 2025-11-12
Date of Manufacture 2024-08-12

UDI# (01)00385480147132(11)240812(17)251112(10)24225039

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	10.00 - 11.20	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

8/22/2024

Date