



**CERTIFICATE OF ANALYSIS**

STP® pHOx Ultra™ / CCX™ Creatinine Sensor  
Lot Number 24222086

Part Number 48595  
Exp Date 2026-08-09  
Date of Manufacture 2024-08-09  
UDI# (01)00385480485951(11)240809(17)260809(10)24222086

|             | <u>Specs</u> | <u>Pass/Fail</u> |
|-------------|--------------|------------------|
| Slope       | ≥ 5.00       | Pass             |
| pHOx Ctl L4 | 0.7 - 1.3    | Pass             |
| pHOx Ctl L5 | 5.2 - 7.2    | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

08/23/2024

Date